


SHEET 2.1

CONSTANT VOLTAGE DRIVE CONDITIONS



FREQUENCY IN HERTZ

ACOUSTICAL

SENSITIVITY

DEVICE WILL PRODUCE THE SPL LISTED BELOW UNDER TEST CONDITIONS DESCRIBED IN TABLE 3. NOMINAL SENSITIVITY AT IKHZ IS dB RELATIVE TO 20µPa. ALL OTHER VALUES IN dB RELATIVE TO THE SENSITIVITY AT IKHz.

FREQUENCY (Hz)	MINIMUM	NOMINAL	MAXIMUM
200	2.0	4.0	6.0
500	0.0	2.0	4.0
1000	-2.0	96.0	2.0
2100 - 3100	+ .0	4.0	7.0
4500	-12.0		

PORT LOCATION: 12S

TABLE I

TOTAL HARMONIC DISTORTION

DEVICE WILL NOT EXCEED TOTAL HARMONIC DISTORTION LEVELS LISTED BELOW.

FREQUENCY (Hz)	AC DRIVE (V rms)	DC BIAS (V)	LIMIT (%)
500	0.221	0	10
870	0.079	0	6
1300	0.079	0	6

TABLE 2

TEST CONDITIONS

TECT CONDITIONS	
NOMINAL SOURCE VOLTAGE	0.079 V rms, 0 mA DC BIAS
SOURCE IMPEDANCE	< I Ohm
TUBING	
COUPLER CAVITY	2 CM ³ , SIMULATED ANSI S3.7 TYPE HA-3 (IEC 126)

TABLE 3

ELECTRICAL

DC RESISTANCE	23 Ohms ± 10%
IMPEDANCE @ 500 Hz	34 Ohms \pm 15%
IMPEDANCE @ IkHz	54 Ohms ±15%

TABLE 4

ISOLATION: CASE WILL BE ELECTRICALLY ISOLATED FROM THE COIL CIRCUIT.

TEMPERATURE: OPERATING RANGE FROM 0°C TO 63°C (SENSITVITY WILL NOT VARY BY MORE THAN ±3 dB WITHIN RANGE)

Revision C.O. # Implementation Date

SENSITIVITY AT 0°C IS 2dB LOWER THAN THE SENSITIVITY AT ROOM TEMPERATURE. DELTA PEAK IS 1.5dB HIGHER AT BODY TEMPERATURE (37°C)}

STORAGE RANGE FROM -40°C TO 63°C

B A	C10109708P C10103482	10-28-10 12-7-05		Active		В
CRITERIA CORRELATION OF TEST FOULDMENT WITH KNOWLES IS ALSO REQUIRED FOR			DR. BY	DATE 12-7-05		
				CK. BY	DATE	
TITLE:	RE	CEIVER		FFH-26786-104	GJP APP. BY	12-8-05 DATE
	PERFORMAN	NCF SPECIFICATION	$_{\rm N}$	SHT 2.1	G.JP	12-8-05

RELEASE LEVEL

REVISION

KNOWLES ELECTRONICS
ITASCA, ILLINOIS U.S.A.

Mouser Electronics

Authorized Distributor

Click to View Pricing, Inventory, Delivery & Lifecycle Information:

Knowles:

FFH-26786-I04