

Certificate



Certificate no.

TA 50226283 01

License Holder:

XP Power Limited
401 Commonwealth Drive,
Haw Par Technocentre,
Lobby B, #02-02, 149598
Singapore

Manufacturing Plant:

ZTW1-BHU-10034600 001

Test report no.: ZTW1-BHU 10034600 001

Client Reference: 8208400537/CL060412

Tested to: IEC 60601-1:2005
EN 60601-1:2006

Certified Product: (Switching Power Supply)

License Fee - Units

Type Designation	: PBM300Pxy (XP)	5
x stands for	: S03, S05, S12, S15, S24, S30, S48, Q01, Q02, Q03, Q04, Q05, Q06, Q07, Q08, Q09, Q10, Q11, Q12, Q13, Q14	1
y stands for	: B or C	1
Rated Voltage	: AC 100-240V, 47-63Hz	
Rated Current	: 4.6-1.6A	
Output Voltage	: see Appendix	
Output Current	: see Appendix	
max. Output Power	: see Appendix	
max. Ambient Temperature	: 50°C	
Protection Class	: I	

This certificate is valid for maximum 5 years.

7

Licensed Test mark:


TÜV Rheinland Taiwan Ltd.

Signature


Date of Issue

(day/mo/yr)

12/04/2012



Dipl.-Ing. W. Hsu



Certificate



Certificate no.

TA 50226283 02

License Holder:

XP Power Limited
401 Commonwealth Drive,
Haw Par Technocentre,
Lobby B, #02-02, 149598
Singapore

Manufacturing Plant:

ZTW1-BHU-10034600 001

Test report no.: ZTW1-SCL 10034600 002

Client Reference: CL-20140826

Tested to: IEC 60601-1:2005
EN 60601-1:2006+A11

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01

Change

Test Requirement : see above

This certificate is valid for maximum 5 years.


Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Dipl.-Ing. W. Hsu



Date of Issue
(day/mo/yr)

26/09/2014

Certificate



Certificate no.

TA 50226283 03

License Holder:

XP Power Limited
401 Commonwealth Drive,
Haw Par Technocentre,
Lobby B, #02-02, 149598
Singapore

Manufacturing Plant:

ZTW1-BHU-10034600 001

Test report no.: ZTW1-SCL 10034600 003

Client Reference: CL-05082015-A12

Tested to: IEC 60601-1:2005
EN 60601-1:2006+A11+A12

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01

Change

Test Requirement : see above

This certificate is valid for maximum 5 years.


Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Dipl.-Ing. W. Hsu



Date of Issue

(day/mo/yr)

30/12/2015

Appendix to TÜV approved Certificate No.: TA 50226283

Certified Product : Switching Power Supply

Report Number : 10034600 001

Type Designation : PBM300Pxy (XP)

(x=S03, S05, S12, S15, S24, S30, S48, Q01, Q02, Q03, Q04, Q05, Q06, Q07, Q08, Q09, Q10, Q11, Q12, Q13, Q14; y=B or C)

Model Name	V1 (V/A max.)	V2 (V/A max.)	V3 (V/A max.)	V4 (V/A max.)	Total Watts max.	
					For y=B (Without Fan, V1 & V2 de-rated to 50%)	For y=C or y=B with additional 35 CFM fan 1)
PBM300PS03y	3.3/60.7	--	--	--	100	200
PBM300PS05y	5.1/60	--	--	--	150	300
PBM300PS12y	12/25	--	--	--	150	300
PBM300PS15y	15/20	--	--	--	150	300
PBM300PS24y	24/12.5	--	--	--	150	300
PBM300PS30y	30/10	--	--	--	150	300
PBM300PS48y	48/6.3	--	--	--	150	300
PBM300PQ01y	3.3/35	5.1/22	12/4	12/4	125	250
PBM300PQ02y	5.1/35	12/10	12/4	5.1/4	150	300
PBM300PQ03y	5.1/35	15/8	15/4	24/2.5	150	300
PBM300PQ04y	5.1/35	12/10	12/4	12/4	150	300
PBM300PQ05y	5.1/35	12/10	12/4	24/2.5	150	300
PBM300PQ06y	5.1/35	12/10	12/4	15/4	150	300
PBM300PQ07y	5.1/35	24/5	12/4	12/4	150	300
PBM300PQ08y	5.1/35	24/5	5.1/4	15/4	150	300
PBM300PQ09y	5.1/35	12/10	5.1/4	24/2.5	150	300



Appendix to TÜV approved Certificate No.: TA 50226283

Certified Product : Switching Power Supply

Report Number : 10034600 001

Type Designation : PBM300Pxy (XP)

(x=S03, S05, S12, S15, S24, S30, S48, Q01, Q02, Q03, Q04, Q05, Q06, Q07, Q08, Q09, Q10, Q11, Q12, Q13, Q14; y=B or C)

PBM300PQ10y	24/6.3	12/10	5.1/4	12/4	150	300
PBM300PQ11y	24/6.3	12/10	5.1/4	24/2.5	150	300
PBM300PQ12y	24/6.3	12/10	12/4	12/4	150	300
PBM300PQ13y	24/6.3	24/5	5.1/4	15/4	150	300
PBM300PQ14y	24/6.3	24/5	12/4	12/4	150	300

Date: 12th. Apr., 2012

Certification Body

W. Hsu
Dipl.-Ing. W. Hsu

